



## Meteorological precision minimum thermometers

acc. to DIN 58653

Capillary: colorless, round

Filling: alcohol, red

Immersion: total

With blue index marker

With metal cap

Suitable for official certification

Minimum thermometers are determined for meteorological measurements with subsequent reading of minimum temperatures and have to be operated in a horizontal position. In case of sinking temperature the index marker below the wetting filling will be pulled downward and remains in this position. In advance of the next measurement the upper part of the thermometer must be tilted downward to ensure that the index marker returns to the bottom of the filling.

Measuring range °C	Scale °C	Total length L <sub>1</sub> mm	Ref. No.
-40 +40	0,5	300	1162109
-30 +50	0,5	300	1162110
-40 +40	0,2	360	1162114
-30 +50	0,2	360	1162115

Further measuring ranges on request.

## Accessory:

Stand for both, a maximum and a minimum thermometer consisting of:

tripod, rod and bracket for two thermometers

**Ref.-No: 205601**

## Additional Order No.

...../01 Officially calibrated without certificate

...../02 Officially calibrated with certificate

...../03 With works certificate

...../04 With DKD calibration certificate